

AMENDMENT OF SOLICITATION/MODIFICATION OF CONTRACT			1. CONTRACT ID CODE	PAGE OF PAGES 1 2
2. AMENDMENT/MODIFICATION NO. 0001	3. EFFECTIVE DATE 30 JUL 1998	4. REQUISITION/PURCHASE REQ. NO.	5. PROJECT NO. (If applicable)	
6. ISSUED BY CODE	N00173	7. ADMINISTERED BY (If other than Item 6)		CODE

CONTRACTING OFFICER
NAVAL RESEARCH LABORATORY
ATTN: CODE 3220.HA
4555 OVERLOOK AVENUE SW,

8. NAME AND ADDRESS OF CONTRACTOR (No., street, country, State and ZIP Code)

TO ALL OFFERORS

✓

9A. AMENDMENT OF SOLICITATION NO.

N00173-98-R-HA04

9B. DATED (SEE ITEM 11)

17 JUL 1998

10A. MODIFICATION OF CONTRACT/ORDER NO.

10B. DATED (SEE ITEM 13)

CODE FACILITY CODE

11. THIS ITEM ONLY APPLIES TO AMENDMENTS OF SOLICITATIONS

☒ The above numbered solicitation is amended as set forth in Item 14. The hour and date specified for receipt of Offers is extended. ☒ is extended, ☐ is not extended.

Offers must acknowledge receipt of this amendment prior to the hour and date specified in the solicitation or as amended, by one of the following methods:

(a) By completing Items 8 and 15, and returning 2 copies of the amendment; (b) By acknowledging receipt of this amendment on each copy of the offer x submitted; or (c) By separate letter or telegram which includes a reference to the solicitation and amendment numbers. FAILURE OF YOUR ACKNOWLEDGMENT TO BE RECEIVED AT THE PLACE DESIGNATED FOR THE RECEIPT OF OFFERS PRIOR TO THE HOUR AND DATE SPECIFIED MAY RESULT IN REJECTION OF YOUR OFFER. If by virtue of this amendment you desire to change an offer already submitted, such change may be made by telegram or letter, provided each telegram or letter makes reference to the solicitation and this amendment, and is received prior to the opening hour and date specified.

12. ACCOUNTING AND APPROPRIATION DATA (If required)

13. THIS ITEM APPLIES ONLY TO MODIFICATIONS OF CONTRACTS/ORDERS, IT MODIFIES THE CONTRACT/ORDER NO. AS DESCRIBED IN ITEM 14.

(✓)	A. THIS CHANGE ORDER IS ISSUED PURSUANT TO: (Specify authority) THE CHANGES SET FORTH IN ITEM 14 ARE MADE IN THE CONTRACT ORDER NO. IN ITEM 10A.
	B. THE ABOVE NUMBERED CONTRACT/ORDER IS MODIFIED TO REFLECT THE ADMINISTRATIVE CHANGES (such as changes in paying office, appropriation date, etc.) SET FORTH IN ITEM 14, PURSUANT TO THE AUTHORITY OF FAR 43.103 (b).
	C. THIS SUPPLEMENTAL AGREEMENT IS ENTERED INTO PURSUANT TO AUTHORITY OF:
	D. OTHER (Specify type of modification and authority)

E. IMPORTANT: Contractor ☐ is not, ☒ is required to sign this document and return 2 copies to the issuing office.

14. DESCRIPTION OF AMENDMENT/MODIFICATION (Organized by UCF section headings, including solicitation/contract subject matter where feasible.)

(SEE PAGE 2)

Except as provided herein, all terms and conditions of the document referenced in Item 9A or 10A, as heretofore changed, remains unchanged and in full force and effect.

15A. NAME AND TITLE OF SIGNER (Type or print)		16A. NAME AND TITLE OF CONTRACTING OFFICER (Type or print)	
15B. CONTRACTOR/OFFEROR (Signature of person authorized to sign)	15C. DATE SIGNED	16B. UNITED STATES OF AMERICA BY (Signature of Contracting Officer)	16C. DATE SIGNED

This amendment is being issued to answer questions that were submitted by prospective offerors. The closing date for the receipt of proposal shall be revised as stated bellow.

1. QUESTION #1: What is the dimension of the chamber?

- **ANSWER:** The chamber dimensions shall be determined by the vendor in order to meet all the specifications outlined in the statement of work (SOW). The final dimensions of the sample and detector stage will depend on the particular dimensions in the sample and detector stage to be used by the vendor. Dictating a priori the chamber dimensions may have restricted in the flexibility of the vendor in addressing the specifications in the SOW. As described in the SOW associated drawing, the chamber is roughly 22" long, 22" high, and 8-10" wide. The only critical dimension is the height of the incident ray port from the floor (the x-ray beam is 43" from the floor).

2. QUESTION #2: Is bracing used? If so, where?

- **ANSWER:** External bracing is used if the vendor deems it necessary to mechanically support the chamber by the stand. Internal bracing is used to insure that the detector and sample arms are free of significant vibrations.

3. QUESTION #3: What is the angle of the 6" CFF evaporation port?

- **ANSWER:** The CFF evaporation port is angled at 45 degrees.

4. The closing date of the solicitation is revised as follows:

FROM: 03 AUG 1998, 4:00 EST
TO: 10 AUG 1998, 4:00 EST